

British Society of Audiology

Promoting excellence in hearing and balance



Recommended Procedure

Pure-tone air-conduction and bone-conduction threshold audiometry with and without masking

Date: August 2018

Due for review: 2023



General foreword

This document presents a Recommended Procedure by the British Society of Audiology (BSA). A Recommended Procedure provides a reference standard for the conduct of an audiological intervention that represents, to the best knowledge of the BSA, the evidence-base and consensus on good practice given the stated methodology and scope of the document and at the time of publication.

Although care has been taken in preparing this information, the BSA does not and cannot guarantee the interpretation and application of it. The BSA cannot be held responsible for any errors or omissions, and the BSA accepts no liability whatsoever for any loss or damage howsoever arising. This document supersedes any previous recommended procedure by the BSA and stands until superseded or withdrawn by the BSA.

Comments on this document are welcomed and should be sent to:

British Society of Audiology
Blackburn House,
Redhouse Road
Seafield,
Bathgate
EH47 7AQ
Tel: +44 (0)118 9660622

bsa@thebsa.org.uk
www.thebsa.org.uk

Published by the British Society of Audiology

© British Society of Audiology, 2018

All rights reserved. This document may be freely reproduced in its entirety for educational and not-for-profit purposes. No other reproduction is allowed without the written permission of the British Society of Audiology.





1. Contents

2.	Introduction	5
3.	Scope	6
3.1	Subject	6
3.2	Procedures	6
4.	Equipment and test environment	6
4.1	Audiometric Equipment	6
4.2	Audiometric Test Environment	6
5.	Preparation for testing	7
5.1	Preparation of test subjects	7
5.2	Test Time	8
6.	Air-conduction audiometry without masking	8
6.1	Instructions	8
6.2	Subjects response	9
6.3	Earphones	9
6.4	Test Order	10
6.5	Timing of test stimuli	10
6.6	Initial familiarisation	10
6.7	Method of finding threshold	11
6.8	Variations in method	11
7.	Bone-conduction audiometry without masking	12
7.1	Bone vibrator	12
7.2	Test frequencies and test order	12
7.3	Test stimuli	13
7.4	Instructions	13
7.5	Methods for finding thresholds	13
7.6	Vibrotactile threshold	13
7.7	Limitations of bone vibrators	13
8.	Cross-hearing and masking	14
8.1	Cross-hearing and its prevention by masking	14
8.2	The principles of masking.....	15
8.3	Masking noise.....	15
8.4	Effective masking level.....	15
8.5	Measuring the thresholds for masking noise (M) if required...	15
8.6	Indicators of cross-masking and rules of masking.....	16
8.6.1	Rule 1.....	16
8.6.2	Rule 2.....	16
8.6.3	Rule 3.....	17
8.7	Instructions for masking.....	17





8.8	Procedure for masking.....	18
8.9	Interpretation of the masking function.....	20
8.9.1	When cross-hearing is not present.....	20
8.9.2	When cross-hearing is present.....	21
8.9.3	Central masking.....	23
8.9.4	Cross-masking.....	24
8.10	Caution.....	26
9.	Audiometric descriptors.....	26
10.	Recommended format for audiogram forms	27
10.1	Audiogram forms.....	27
10.2	Symbols.....	28
10.3	Working Audiograms.....	30
10.4	Masking Levels.....	30
10.5	Notes.....	31
11.	Calibration.....	31
11.1	Stage A: routine checking and subjective tests.....	31
11.2	Stage B: periodic objective tests.....	33
11.3	Stage C: basic calibration tests.....	33
12.	References.....	34
	Appendix A. Authors and acknowledgments.....	36
	Appendix B. Standards relevant to audiometry.....	36
	Appendix C. Permitted ambient noise levels for audiometry.....	37





2.1 Introduction

2 This document replaces the previous version (BSA 2011). Its purpose is to describe
3 standard procedure and recommendations for effective pure-tone audiometry carried
4 out in most audiological contexts. It also includes descriptors for pure-tone audiograms
5 and the recommended format for audiogram forms. This document is not intended to
6 provide guidance on specific circumstances or on interpretation of results. It is
7 important that the competent person carrying out, or responsible for, the test (the
8 ‘tester’) uses professional judgement when deciding on the particular approach to be
9 used with each person being tested (the ‘subject’), given the specific circumstances and
10 the purposes of the test, and the tester’s level of competency. The BSA has produced a
11 separate procedure for the determination of uncomfortable loudness levels (BSA).

12 The term ‘shall’ is used in this document to refer to essential practice, and ‘should’ to
13 refer to desirable practice.

14 Unless stated otherwise, this document represents the consensus of expert opinion and
15 evidence as interpreted by the Professional Guidance Group of the BSA in consultation
16 with its stakeholders (Appendix A). The document was developed in accordance with
17 the BSA Procedure for Processing Documents (BSA).

18

19 Shared Decision-making

20 It is implied throughout this document that the service user should be involved in
21 shared decision-making when undertaking audiological intervention, receiving
22 subsequent information and understanding how it will impact on the personalisation of
23 care. Individual preferences should be taken into account and the role of the clinician is
24 to enable a person to make a meaningful and informed choice. Audiological
25 interventions bring a variety of information for both the clinician and the patient which
26 can be used for counselling and decision-making regarding technology and anticipated
27 outcomes.

28

29





330 Scope

331 Subjects

32 This document describes procedures suitable for routine clinical use with adults and
33 older children. It may not be appropriate for certain populations (e.g. adults with
34 intellectual disabilities and younger children). In these cases some modification of the
35 test method may be required, although this may result in a less accurate measurement
36 of hearing threshold levels.

332 Procedures

38 Procedures are described for manual pure-tone audiometry, using both air-conduction
39 (a-c) and bone-conduction (b-c) testing, with and without masking. The document does
40 not cover high-frequency audiometry (>8000 Hz), screening audiometry, use of short-
41 duration tone bursts, self-recording audiometry or sound-field audiometry.

442 Equipment and test environment

443 Audiometric equipment

44 The audiometer, transducers and response button shall be clean. Audiometers shall
45 meet the performance and calibration requirements of the relevant and current BS EN
46 ISO standards (see Section 11 and Appendix B).

442 Audiometric test environment

48 The subject's face shall be clearly visible to the tester. The subject shall not be able to
49 see or hear the tester adjust the audiometer controls. When the test is controlled from
50 outside the audiometric test room, the subject shall be monitored through a window or
51 by a closed-circuit TV system. Audible communication with the subject should also be
52 possible.

53 Excessive ambient noise will affect the test results, and ambient noise should not exceed
54 the levels set out in the BS EN ISO standards (see Appendix C for further comments and
55 details of the permitted ambient noise). The problems caused by ambient noise are
56 greater when testing by b-c as there are no earphones in place to reduce the noise





57 reaching the ears. Testers should also be alert to the problems of intermittent or
58 transient noise during the test.

59 In general, the ambient noise should not exceed 35 dB (A) as measured with a calibrated
60 sound level meter. If it is higher than this then it is recommended that audiometry
61 should not proceed.

562 Preparation for testing

563 Preparation of test subjects

64 The tester shall adopt an effective communication strategy with the subject throughout.
65 This must take account of the subject's age, hearing, language skills and any other
66 possible communication difficulties. Any significant communication problems shall be
67 recorded as these may affect the subject's performance.

68 Audiometry shall be preceded by otoscopic examination (see Recommended Procedure:
69 Ear Examination (BSA)) and the findings recorded, including the presence of wax.

70 Occluding wax may be removed prior to audiometry but if wax is removed the
71 procedure shall only be undertaken by someone who is qualified and competent to do
72 so.

73 If there is a likelihood of ear canals collapsing with supra-aural earphones in position this
74 shall be recorded as it may lead to measurement of a false air-bone gap. In some cases
75 the use of insert earphones (e.g. Etymotic ER3 and ER5) will avoid this problem (see
76 Section 6.3).

77 The subject shall be asked about any exposure to loud noise during the previous 24
78 hours, as this can cause a temporary hearing loss. If the answer is yes then more details
79 should be obtained regarding the exposure and results recorded. 'Loud' can be
80 determined by having to shout or use a raised voice to communicate at a distance of 1
81 metre or 3 feet. If the results may have been affected by recent noise exposure then it
82 may be necessary to re-test the subject at a time when they have had no recent
83 exposure to noise.

84 Subjects shall be asked if they have tinnitus, as this may affect their ability to detect
85 tones in one or both ears (see Section 6.8). Subjects shall be asked if they have better





86 hearing in one ear; if so testing should commence with that ear, otherwise testing can
87 start in either ear.

88 If applicable, inform the subject about intercom facilities. After giving the test
89 instructions, remove any hearing aids, also any glasses, headwear or earrings that may
90 obstruct the correct placement of the transducers, cause discomfort or affect sound
91 transmission. Wherever possible, hair, scarves etc, should not be allowed to sit between
92 the ear and the transducer.

592 Test time

94 Care should be taken not to fatigue the subject as this can affect the reliability of the
95 test results. If the test time exceeds 20 minutes, subjects may benefit from a short
96 break.

697 Air-conduction audiometry without masking

698 Instructions

99 Instructions shall give clear information about the task. This could be as follows:

100 "I am going to test your hearing by measuring the quietest sounds that you can hear. As
101 soon as you hear a sound (tone), press the button. Keep it pressed for as long as you
102 hear the sound (tone), no matter which ear you hear it in. Release the button as soon as
103 you no longer hear the sound (tone). Whatever the sound, and no matter how faint the
104 sound, press the button as soon as you think you hear it, and release it as soon as you
105 think it stops."

106 Alternative wording is acceptable providing the same points of instruction are included.
107 The provision of an abbreviated printed version of these instructions may be
108 advantageous. The subject should be asked if they understand the instructions. They
109 should also be told that they should sit quietly during the procedure and may interrupt
110 the testing in case of discomfort.

111 Subjects with tinnitus present at the time of the test should be asked to ignore their
112 tinnitus as much as possible and to respond to the test tones. They should be instructed
113 to inform the tester if they experience difficulty in discriminating between their tinnitus





114 and the test tones. A note to that effect should be made on the audiogram form,
115 including which frequencies were affected (see also Section 6.8).

612 Subject's response

117 The subject's response to the test tone should clearly indicate when the test tone is
118 heard and when it is no longer heard. The response system should be inaudible, with a
119 response button connected to a signal light the preferred method. When testing
120 younger children, adults with intellectual disabilities or subjects with attention
121 difficulties a more engaging response method may be required, and, if so, this shall be
122 recorded.

623 Earphones

124 There are three main types of transducers that can be used for air-conduction
125 audiometry: supra-aural, circum-aural and insert earphones. Supra-aural earphones (e.g.
126 Telephonics TDH39 and TDH49) rest on the ear and have traditionally been used for a-c
127 audiometry. Circum-aural earphones (Sennheiser HDA200) surround and cover the
128 entire ear. However, both supra- and circum-aural earphones can be cumbersome,
129 particularly when used for masking bone-conduction thresholds, and may cause the ear
130 canal to collapse. Insert earphones (e.g. Etymotic Research ER3 and ER5) use a
131 disposable foam tip for directing the sound straight into the ear canal and therefore
132 prevent the ear canal from collapsing. Insert earphones are also associated with less
133 transcranial transmission of sound than supra-aural earphones so reduce the need for
134 masking (see Section 8.1). However, insert earphones may not be appropriate in ears
135 with infections, obstructions or abnormalities. In cases of excessive wax, insert
136 earphones could also push the wax further into the canal and therefore must be
137 avoided.

138 The tester shall fit the earphones and the subject should be instructed not to hold or
139 move them, after checking with the subject that there is no discomfort. The sound
140 opening of a supra- or circum-aural earphone shall be aligned with the ear canal
141 entrance. If insert earphones are used, the appropriately sized ear tip of an insert
142 earphone should be inserted so the outer end is flush with the entrance to the ear canal.
143 In all cases, incorrect placement may invalidate calibration and provide less protection
144 from ambient noise.





145 With a-c, vibrotactile perception can occur at frequencies of 500 Hz and below, and at
146 high hearing levels. The tester should be aware of the possibility that thresholds at these
147 frequencies and levels may be vibrotactile.

648 Test order

149 Start with the better-hearing ear (according to the subject's account) and at 1000 Hz.
150 Next, test 2000 Hz, 4000 Hz, 8000 Hz, 500 Hz and 250 Hz in that order. Then, for the first
151 ear only, retest at 1000 Hz. If the retest value is no more than 5 dB different from the
152 original value take the more sensitive threshold as the final value, but if the retest value
153 differs from the original value by more than 5 dB then the reason for the variation shall
154 be investigated. The subject may need to be re-instructed and the full test repeated for
155 that ear (but see also Sections 5.2 and 6.8 regarding the effects of a test taking too
156 long). Unusually variable results shall be noted on the audiogram. Where needed and
157 practicable, test also at intermediate frequencies 750 Hz, 1500 Hz, 3000 Hz and 6000 Hz
158 (3000 Hz and 6000 Hz may be required in cases of high-frequency hearing loss). Test the
159 opposite ear in the same order. The retest at 1000 Hz is normally not required in the
160 second ear unless tests in the first ear revealed significant variation.

665 Timing of the test stimuli

162 The duration of the presented tone shall be varied between 1 and 3 seconds. The
163 interval between the tones shall be varied between 1 second and at least 3 seconds. The
164 tester must ensure that the timing of each tone is not predictable; random variations in
165 durations are intended as a check against false positive responses. It is important that
166 the tester does not stop the signal as soon as the subject responds, signals must be of
167 the full duration and the subject must respond throughout each one.

668 Initial familiarisation

169 To ensure the subject is familiar with the task, present a tone of 1000 Hz that is clearly
170 audible (e.g. at 40 dB HL for a normally hearing subject or approximately 30 dB above
171 the estimated threshold for a subject with a hearing impairment, but never more than
172 80 dB HL). If there is no response, increase the level of the tone in 10-dB steps until a
173 response occurs. If the tone is still inaudible at 80 dB HL, increase the level of the tone in
174 5-dB steps until a response occurs, taking care to monitor the subject for discomfort.





175 If the responses are consistent with the tone presentation (i.e. onset and offset) the
176 subject is familiarised with the task. If not, repeat. If after this repeat the responses are
177 unsatisfactory, re-instruct the subject.

678 Method for finding threshold

- 179 1. Following a satisfactory positive response, reduce the level of the tone in 10-dB
180 steps until no further response occurs.
- 181 2. Increase the level of the tone in 5-dB steps until a response occurs.
- 182 3. After the first response using an ascending approach, decrease the level by 10 dB
183 and begin another ascending 5-dB series until the subject responds again.
- 184 4. Continue to decrease the level by 10 dB and increase by 5 dB until the subject
185 responds at the same level on two out of two, three or four (i.e. 50 % or more)
186 responses on the ascent. This is the hearing threshold level. Threshold is defined as
187 the lowest level at which responses occur in at least half of a series of ascending
188 trials with a minimum of two responses required at that level.
- 189 5. Proceed to the next frequency, starting at a clearly audible level (e.g. 30 dB above
190 the adjacent threshold, but see notes on familiarisation in Section 6.6) and use the
191 10-dB-down, 5-dB-up sequence described in Step 4 until the threshold criterion is
192 satisfied.

698 Variations in method

194 There will be situations where the test frequencies will vary from those in Section 6.4.
195 For example, industrial audiometry (Health & Safety Executive, 2005, Appendix 5)
196 requires testing at 500 Hz, 1000 Hz, 2000 Hz, 3000 Hz, 4000 Hz, 6000 Hz and 8000 Hz.

197 Subjects with short attention spans, and some elderly subjects, may find the full test
198 rather tiring. In these cases it may be appropriate to test fewer frequencies, as it is
199 better to test fewer frequencies accurately than to attempt a complete test on an
200 uncooperative subject where the accuracy will be in doubt. When frequencies are
201 omitted from the test the reason shall be recorded. With such subjects other
202 modifications to technique may be required, such as the use of longer test tones or
203 alternative response methods. Again, these variations in technique shall be recorded.





204 If the subject is unable to perform an accurate test at some frequencies due to an
205 inability to distinguish between their tinnitus and the test tone, then a frequency-
206 modulated or warble tone may be used as a stimulus. Subjects often find a warble tone
207 easier to hear against their tinnitus than a pure-tone. However, there are calibration and
208 perception problems with warble tones, and the thresholds measured may be in error
209 as a result. It is essential to record at which frequencies and in which ears warble tones
210 were used. The use of pulsed tones and narrow-band noise is not advised due to
211 calibration and perception problems.

212 If there is reason to suspect that the hearing thresholds are exaggerated, published
213 variations in technique may help with this (e.g. Cooper & Lightfoot, 2000). When used, a
214 note citing the method shall be added to the audiogram form (Section 10).

7.1.5 Bone-conduction audiometry without masking

216 Without masking, it is not possible to determine which ear is responding to bone-
217 conduction testing. Calibration standards for b-c apply only to monaural hearing and
218 were derived using approximately 35 dB sensation level of masking noise in the non-test
219 ear. When testing without masking, thresholds may appear more acute by about 5 dB
220 due to binaural stimulation.

7.2.1 Bone vibrator

222 The bone vibrator is normally initially placed over the mastoid prominence of the worse
223 hearing ear (as defined by the a-c thresholds averaged between 500 Hz and 2000 Hz),
224 with the required area of the vibrator in contact with the skull. It shall be placed as near
225 as possible behind the pinna without touching it and without resting on hair. The
226 vibrator shall be held firmly in place by means of a headband that holds it against the
227 skull with the required static force. The side on which the vibrator was placed shall be
228 noted on the audiogram form. See Section 10.2 for details of the use of symbols. An
229 alternative placement of the bone vibrator is on the forehead but this site requires a set
230 of correction values, which are available in BS EN ISO 389-3.

7.3.2 Test frequencies and test order

232 The preferred test order is similar to that used in a-c audiometry starting at 1000 Hz,
233 however alternatives are acceptable. Bone-conduction should normally only be





234 performed in the frequency range 500 Hz to 2000 Hz, and it may not be necessary or
235 appropriate to test at all these frequencies in every case. No retest is required at 1000
236 Hz. See Section 7.7 for comments on limitations of bone vibrators and for test
237 frequencies outside this range.

738 Test stimuli

239 The requirements for timing of the test stimuli are the same as for a-c audiometry. See
240 Section 6.5. The initial presentation level should be clearly audible to the subject (with
241 reference to the measured a-c thresholds at each frequency).

744 Instructions

243 Instructions are the same as for a-c audiometry, as described in Section 6.1. However,
244 emphasis should be given that the subject should respond regardless of the side on
245 which the sound (tone) is heard.

745 Methods for finding threshold

247 Ear-specific b-c audiometry requires masking of the non-test ear. Where an ear-specific
248 measure is not required, b-c audiometry may be undertaken without masking. The ear
249 being tested by b-c should not be occluded.

250 Determine hearing threshold levels as described in Section 6.7.

756 Vibrotactile threshold

252 For mastoid location of the bone vibrator, vibrotactile threshold may be as low as 25 dB
253 at 250 Hz, 55 dB at 500 Hz and 70 dB at 1000 Hz (Boothroyd and Cawkwell, 1970).
254 However, there is large inter-subject variation in vibrotactile thresholds. Care must be
255 taken not to misinterpret vibrotactile perceptions as hearing. Any threshold considered
256 vibrotactile shall be noted on the audiogram form.

757 Limitations of bone vibrators

258 The standard bone vibrator used in audiometry (Radioear B71) has poor distortion
259 performance at low frequencies (Lightfoot, 2000). Testing is not recommended at
260 frequencies below 500 Hz because the subject's threshold may relate to hearing at the





261 second or third harmonic rather than the fundamental. Bone-conduction tests above
262 2000 Hz are also problematic due to transducer limitations and calibration issues, and
263 should be avoided (Lightfoot and Hughes, 1993, Margolis et al 2013). However, there
264 may be exceptional circumstances when tests at the lower and higher frequencies are
265 required, depending on the investigation performed. A check must be made that these
266 frequencies have been included in periodic objective calibration tests, and caution is
267 advised in the interpretation of the results.

268 Headband tension has an impact on the sound levels delivered. It is difficult to measure
269 the actual headband tension *in situ*, but testers need to be aware of this source of error
270 (e.g. with a small head) and record any suspected errors from this source.

871 **Cross-hearing and masking**

872 **Cross-hearing and its prevention by masking**

273 Although earphones allow sound to be presented to one ear at a time, it is not always
274 certain that the intended (test) ear is the one actually detecting the sound. When the
275 hearing acuity of the ears is very different it is possible that, when testing the worse ear,
276 the better (non-test) ear detects the test signals more easily despite the fact that the
277 signals reaching it are attenuated.

278 This interaural attenuation, also referred to as transcranial transmission loss, varies
279 considerably from person to person. It is also earphone dependent. It varies between
280 40-80 dB when using supra-aural or circum-aural earphones. When using insert
281 earphones, the transcranial transmission loss is higher, with a minimum transcranial
282 transmission loss of 55 dB if the earphones are inserted correctly (Munro and Agnew,
283 1999). With b-c testing there can be little or no transcranial transmission loss.

284 When the difference in the thresholds of the two ears is greater than the transcranial
285 transmission loss, cross-hearing may occur and the apparent threshold of the worse ear
286 is in fact a 'shadow' of the better ear.

287 Reliance should not be placed on the subject to make an accurate report of the ear in
288 which the sounds were heard, since many people are unable to make such judgements
289 easily and the sound may not be fully lateralised to one ear.





890 The principles of masking

291 The problems of cross-hearing can usually be overcome by temporarily elevating the
292 hearing threshold of the non-test ear by a known amount to enable an accurate
293 assessment of the test ear threshold to be made. This is achieved by presenting a
294 masking noise into the non-test ear at the appropriate intensity to prevent it from
295 detecting the test signals, and at the same time measuring the apparent threshold of
296 the test ear with the test signals. There is normally a 1:1 relationship between the
297 increase in masking noise and the elevation of the masked threshold of the non-test ear.

298 The term 'not-masked' is used to describe measurements made without masking, rather
299 than the term 'unmasked' which refers to different psychophysical phenomena.

800 Masking noise

301 Narrow-band masking noise of the type specified in BS EN ISO 389-4 should be used,
302 where the geometric centre frequency coincides with that of the test tone and the
303 bandwidth of the noise is between one-third and one-half of an octave.

804 Effective masking level

305 Masking noise should be calibrated in terms of effective masking level (EML) according
306 to BS EN ISO 389-4. In the presence of masking noise at a particular centre frequency
307 and effective masking level, the pure-tone threshold of hearing (dB HL) at that
308 frequency will be raised to that level. For example, a 1000 Hz noise at 50 dB EML
309 presented to an ear will normally raise its hearing threshold for a 1000 Hz pure-tone to
310 50 dB HL.

311 When masking noise is calibrated in terms of effective masking level it is not necessary
312 to measure the subject's hearing threshold for the masking noise prior to testing with
313 masking (see Sections 8.5 and 8.8).

815 Measuring the threshold for masking noise (M) if required

315 When the noise is not calibrated in effective masking level, the threshold for masking
316 noise (M) shall be measured. This indicates the lowest level of a masking noise that can
317 be detected, measured in dB (relative to an arbitrary zero). The initial masking level
318 used shall be M+10 (see Section 8.8).





319 If possible the same method as used for pure-tone threshold determination should be
320 used to determine M. The procedure should be repeated for each masking noise
321 corresponding to the frequencies of the pure-tones to be masked. Where it is not
322 possible to determine M using the usual threshold technique, perhaps because of
323 audiometer design, care should be taken to determine M as accurately as possible.

826 Indicators of cross-hearing and the rules for masking

325 The indicators (or 'rules') given below are to be considered independently at each
326 frequency. Note that words such as 'better' and 'worse' etc. describe hearing as
327 measured by air-conduction. The test ear is always the ear whose hearing threshold is
328 being sought; it is the ear being presented with the pure tone directly. The non-test ear
329 is the ear which may have to be masked to prevent detection of the pure tone.

330 It is preferable to mask two or three frequencies properly, rather than incorrectly or
331 hurriedly masking more frequencies. It is not essential to mask in the order that the
332 rules are given below.

835.1 Rule 1

334 Masking is needed at any frequency where the difference between the left and right
335 not-masked a-c thresholds is 40 dB or more when using supra- or circum-aural
336 earphones or 55 dB or when using insert earphones.

837.2 Rule 2

338 Masking is needed at any frequency where the not-masked b-c threshold is better than
339 the air-conduction threshold of either ear by 10 dB or more. The worse ear (by air-
340 conduction) would then be the test ear and the better ear would be the non-test ear to
341 be masked.

342 *Notes on Rule 2:*

343 Although this rule may frequently indicate the need for masking, there will be occasions
344 where this is not warranted, depending on the purpose of the investigation. For
345 example it may not benefit patient management to mask more than two b-c frequencies
346 on one ear, or to mask small air-bone gaps.





347 If the b-c threshold with masking is more than 10 dB worse than the not masked b-c
348 threshold, then the not-masked b-c threshold can be attributed to the other (non-test)
349 ear. However if the b-c threshold with masking is not more than 10 dB worse than the
350 not masked b-c threshold then it is possible that the not-masked b-c result was from the
351 test ear, and it may be necessary to test the b-c thresholds of the other ear with
352 masking.

853 Rule 3

354 Masking will be needed additionally where Rule 1 has not been applied, but where the
355 b-c threshold of one ear is more acute by 40 dB or more (if supra or circum-aural
356 earphones have been used) or 55 dB or more (if insert earphones have been used) than
357 the not-masked a-c threshold attributed to the other ear.

358 *Notes on Rule 3:*

359 Rule 3 is necessary because an a-c frequency that does not require masking under Rule
360 1, may need to be masked if the b-c results show that the non-test ear has a conductive
361 element. Note that it is the sensitivity of the non-test cochlea (as indicated by the b-c
362 threshold) that is the important factor in cross-hearing, and that Rule 1 is merely a
363 convenient way of anticipating the need to mask in many cases.

364 At frequencies where no b-c thresholds have been measured, doubt may exist regarding
365 the possible effect of Rule 3. If there is a possibility that a-c thresholds at these
366 frequencies (including 250 Hz and 8000 Hz) are not the true thresholds, they should be
367 masked or marked accordingly on the audiogram form.

868 Instructions for masking

369 Suitable instructions would be:

370 “In this next test, you will hear the sounds (tones) again, just as before. I would like you
371 to press the button as soon as you hear the sound (tone) start and release it as soon as
372 it disappears. Do this even for the very faint sounds (tones), and no matter which side
373 you seem to hear the sounds (tones).





374 For some of the time, you will also hear a steady rushing noise, but I want you to ignore
375 it and press the button only when you hear the sounds (tones). This steady rushing noise
376 will get louder at times.

377 I want you to tell me if any of the sounds become uncomfortably loud, or if you would
378 like me to explain the test again.”

379 The subject must not be told to expect to hear the pure-tone in the test ear. The very
380 fact that masking noise is required means that it is not known which ear is picking up the
381 signals.

888 Procedure for masking

383 This procedure is called the plateau-seeking method for masking. It is appropriate for
384 both a-c and b-c testing.

385

386 ***Please note that the term ‘re-establish’ refers to the technique described in 6.6 and 6.7***
387 ***of this document.***

388

389 1. Re-establish hearing threshold in the test ear by presenting the tone and seeking the
390 response without masking noise to remind the subject what to listen for. This is
391 always necessary for b-c because the occluded not-masked hearing threshold level is
392 required.

393 2. Introduce masking noise to the non-test ear. The initial level of masking should be
394 the effective masking level which is equal to the tonal threshold level of that ear at
395 that frequency. Wait a few seconds in case the subject mistakenly responds to the
396 introduction of the masking noise (a response at this stage may require brief re-
397 instruction).

398 3. Re-establish the hearing threshold in the test ear in the presence of masking noise.
399 Take this tone level as the pure-tone threshold at that level of masking.

400 4. Increase the level of masking noise by 10 dB. Re-establish the hearing threshold level
401 in the test ear. Take this tone level as the pure-tone threshold at that level of
402 masking.





- 403 5. Continue to repeat Step 4, using increments of 10 dB in masking noise, until you
404 have at least four measurements (including the initial starting point) and until three
405 successive measurements yield the same tonal threshold. You may be unable to
406 obtain this plateau due to maximum level of the audiometer being reached or
407 because the subject finds the masking noise uncomfortable; in either case mark as
408 unreached. (See also Sections 8.10 and 8.11).
- 409 6. When three successive levels of masking yield the same tonal threshold, or one
410 threshold is no more than 5 dB different from the other two, this is the 'plateau' (see
411 Figures 1 and 2, and Section 8.10). The mode (i.e. the threshold which occurs 2/3) of
412 the three hearing threshold levels at plateau is taken as the correct hearing
413 threshold of the test ear and no further masking is required. Withdraw the masking
414 noise and plot the hearing threshold level on the audiogram.

415 The use of a masking chart to plot the relationship between the masking noise level and
416 pure-tone threshold can be helpful for interpreting difficult cases. Both axes of the
417 masking chart are marked in dB and the aspect ratio is 1:1. See Figure 1 for an example.

418 **Notes on masking method:**

419 It may be appropriate on occasions to use smaller step sizes when increasing the
420 masking noise, particularly where the plateau is not well defined (see also Sections
421 8.9.4).

422 Some testers use alternative techniques to determine the masked thresholds.
423 Techniques other than those described here are not recommended.

424 Proceed with caution when using masking noise greater than 80 dB EML or tones
425 greater than 100 dB HL (see Step 5 above and Section 8.10).

426 **Masking during bone-conduction testing**

427 An insert earphone should be used to deliver masking noise to the non-test ear for b-c
428 testing, for subject comfort and for the advantages of high transcranial transmission
429 loss. If the insert earphone is not of the type Etymotic ER3 or ER5 or has not been
430 calibrated to effective masking level, then it will be necessary to measure the threshold
431 of masking (M; Section 8.5). A supra-aural or circum-aural earphone can be used if there
432 is no alternative.





433 Step 1 of the masking function (which involves re-establishing the not-masked tonal
434 threshold, but with the non-test ear occluded by an insert earphone or headphone, see
435 Section 8.8) may lead to an improvement of the measured threshold. This is due to the
436 occlusion effect which is more pronounced at the lower frequencies. If an improvement
437 in threshold is noted, the original not-masked threshold value on the audiogram should
438 not be altered although the new value should be used on the masking chart.

439

840 Interpretation of the masking function


441 In the interpretation of the masking results, it is important to remember that all
442 threshold measurements are associated with a degree of uncertainty (at least ± 5 dB).
443 Consequently, the measured masking function may not exactly match the idealised
444 pattern and a 'best fit' approach should be adopted. The following sections provide
445 guidance on the interpretation of the idealised masking functions.

8461 When cross-hearing is not present

447 This is when the original not-masked threshold measurement represents the true
448 threshold of the test ear, even though there was a risk of cross-hearing. It is most often
449 manifest by the measured tone thresholds at at least three masking levels being within
450 5 dB of the not-masked tonal threshold. An example is shown in Figure 1.

451





Test details may be recorded as follows:

Test frequency: 2000 Hz

Test ear: R / L

Test mode: a-c / ~~b-e~~

Pure-tone threshold = 55 dB HL

452

453 *Figure 1 - Recommended design of masking chart and example of masking function*
454 *derived in a case where cross-hearing is not present. The horizontal arrow by the y axis*
455 *represents the not-masked hearing threshold level.*

456

457

458


459

460 **When cross-hearing is present**

461 Cross-hearing occurs when the original not-masked threshold was a 'shadow' of the
462 non-test ear, with the test ear threshold being at a higher level. A typical masking
463 function is illustrated in Figure 2 and usually takes the form of a short (and sometimes
464 absent) initial horizontal line originating from the not-masked threshold (a), followed by
465 a sloping section (b), and then by a horizontal section, or plateau, (c).

466





Test frequency: 500 Hz

Test ear: R / €

Test mode: $\text{a-€} / \text{-b-c}$

Pure-tone threshold = 40 dB HL

467

468 **Figure 2 - Recommended design of masking chart and example of masking function**
469 **illustrating cross-hearing**

470 In Figure 2:

471 (a) represents an initial condition where the masking noise though audible does not
472 have a masking effect. Low masking levels (up to 10 dB above the initial masking
473 level) are typical for this condition. Both tone and masking noise are heard in the
474 non-test ear.

475 (b) represents direct peripheral masking where the threshold of the non-test ear is
476 being raised by the presence of the noise but not enough to prevent it from
477 detecting tones more easily than the test ear. Again, both tone and masking noise
478 are heard in the non-test ear. Note that the slope of this part of the function is
479 always approximately 1 dB per dB (i.e. approximately 45 degrees assuming the
480 recommended chart with aspect ratio 1:1). In cases where this 1 dB per dB slope
481 continues to the audiometer's tonal or masking maximum output limit, the true test
482 ear threshold has not been found and the appropriate audiometric symbol with a
483 downward pointing arrow should be drawn on the audiogram at the last employed
484 (highest) pure-tone intensity (see Figure 6 below).






485 (c) represents the true threshold of the test ear (40 dB HL in this example). At these
486 levels the masking noise has raised the threshold of the non-test ear to the extent
487 that the intensity of the test tone is sufficient to be just audible in this test ear. Note
488 that the function is horizontal at this point: the plateau. At the start of the plateau
489 the subject may hear the tones in the test ear for the first time or may hear the
490 tones centrally and will sometimes report this to the tester. At higher masking
491 intensities, the subject should hear the tone and masking noise in the ears to which
492 they are being presented. The maximum level of masking required to define the
493 plateau could be recorded for training or audit purposes (e.g. in Figure 2 this is 95 dB
494 EML).

3.3 Central masking

496 This refers to the inability of the brain to identify a tone in the presence of masking,
497 even when they are heard in opposite ears, hence masking is occurring centrally rather
498 than peripherally (in the cochlea). This effect is most commonly apparent at the higher
499 masking levels and may be evident as an upward slope of the masking function of
500 consistently less than 1 dB per dB (i.e. between 5 and 35 degrees) which may lead to an
501 inability to determine the plateau. See Figure 3.



Test frequency: 500 Hz
Test ear: R / \downarrow
Test mode: a-c / b-c

Pure-tone threshold = 40 dB HL (approx.)

502
503

Figure 3 - Illustration of central masking





504 Line (e) in Figure 3 is an example of central masking. If a 5-dB increase in threshold is
505 seen at the third point of an otherwise possible plateau, it is wise to go on to mask at
506 higher levels in order to evaluate the slope and so aid interpretation of the masking
507 function. In these cases a reasonable estimate of the true threshold can be made from
508 the masking chart because cross-hearing has been ruled out. In figure 3 the true
509 threshold of 40 db HL is approximated, and this should be clearly indicated on the
510 audiogram.

8.1.4 Cross-masking


512 When sufficient masking has been applied the threshold of the non-test ear is elevated
513 so that cross-hearing cannot occur. In Figure 2 above, effective masking from 75 dB EML
514 has prevented cross-hearing and allowed the true threshold of the test ear to be
515 established, indicated by the plateau (c). However, if at some stage the masking level
516 becomes sufficiently high, it may be capable of providing a masking effect in the *test ear*
517 through transcranial transmission. This is known as cross-masking.

518 Since cross-masking is of peripheral origin, this will be evident as a second
519 approximately 1 dB per dB slope (approximately 45 degrees) on the masking function, as
520 illustrated by line (d) in Figure 4a. Even though the masking noise is reaching the test
521 ear, the subject will only be aware of the tone in that ear since the noise will be much
522 louder in the non-test ear.

523 Cross-masking is primarily through bone-conduction (as with cross-hearing; see notes on
524 Rule 3 in Section 8.6.3) and the point at which the masking signal is detected by the test
525 ear will depend on the bone-conduction hearing of that side.

526 Cross-masking will be a particular problem when the test ear has a conductive loss (with
527 good bone-conduction) and the non-test ear has at least a moderate loss. In this
528 situation high effective masking levels will be required in the non-test ear, which may
529 readily stimulate the cochlea in the test ear leading to cross-masking.





Test frequency: 2000 Hz

Test ear: R / L

Test mode: ~~a-c~~ / b-c

Pure-tone threshold = 40 dB HL

530

531 **Figure 4a - Illustration of cross-masking**

532


533 In some cases the plateau will be shorter than the three points required to indicate the
534 true masked threshold, due to cross-masking occurring. In Figure 4b the plateau is only
535 defined by two points but this may be sufficient to define the true masked threshold.

536 In cases when the potential for cross-masking is apparent, increasing the masking level
537 in 5-dB steps, rather than 10-dB steps might help to identify a shortened plateau.

538 It may be impossible to accurately mask a conductive hearing loss if the plateau is not
539 well defined. Where the masking test could not be performed accurately, or the results
540 are in doubt, this should be clearly indicated on the audiogram. Remember that all
541 threshold measurements are associated with a degree of uncertainty and real-life
542 masking functions will require a 'best fit' approach.

543 There is less risk of cross-masking with insert earphones.





Test frequency: 1000 Hz

Test ear: R / \neq

Test mode: ~~a-c~~ / b-c

Pure-tone threshold = 40 dB HL

544

545 *Figure 4b - Illustration of a masked threshold defined by only two points*

546

8410 Caution

548 Care needs to be taken when using high levels of masking, particularly when testing at
549 several frequencies, as it can present a risk to the subject (see The Control of Noise at
550 Work Regulations, Health & Safety Executive, 2005).

551 In subjects with tinnitus, extra care should be taken when using high levels of masking
552 noise, as this can exacerbate the tinnitus. In some cases, it may be appropriate not to
553 perform masking.

954 Audiometric descriptors

555 The hearing threshold levels of an individual ear are often described in general terms
556 rather than in terms of the actual numbers at different frequencies on a pure-tone
557 audiogram. Recommendations are made below to associate particular descriptors with
558 bands of average hearing impairment.





559 Four audiometric descriptors are given. These are based on the average of the pure-
560 tone a-c hearing threshold levels at 250, 500, 1000, 2000 and 4000 Hz. Averages do not
561 imply any particular configuration of hearing loss and do not exclude additional terms
562 (e.g. profound high-frequency hearing loss) being used.

563	Descriptor	Average hearing threshold levels (dB HL)
564	Mild hearing loss	21-40
565	Moderate hearing loss	41-70
566	Severe hearing loss	71-95
567	Profound hearing loss	In excess of 95

568 For the purposes of this document, in determining the five-frequency average value of
569 hearing loss, if at any frequency no response is obtained due to the severity of the loss,
570 this reading shall be given a value of 130 dB HL.

571 Anomalies may occur in calculating the average hearing loss if an audiometer with
572 insufficient output is used in the measurement of severe and profound hearing loss.

573 **Note:**

574 While audiometric descriptors may provide a useful summary of an individual's hearing
575 thresholds, they shall not be used as the sole determinant for the provision of hearing
576 support. The ability to detect pure tones using earphones in a quiet environment is not
577 in itself a reliable indicator of hearing disability and audiometric descriptors alone shall
578 not be used as the measure of difficulty experienced with communication in background
579 noise, the primary complaint of individuals with hearing loss.

10. Recommended format for audiogram forms

10.1 Audiogram form

582 Hearing threshold level can be plotted graphically on an audiogram form. The aspect
583 ratio should be fixed at 20 dB:1 octave in all cases to ease interpretation. The





584 recommended format is shown in Figure 5. Other audiogram forms are acceptable, as
585 long as the information shown in Figure 5 is recorded.

1862 Symbols

587 Symbols are shown in Figure 5. Air-conduction symbols should be connected with
588 continuous straight lines; bone-conduction symbols should be joined with broken lines.

589 For not-masked bone-conduction, the mastoid on which the bone vibrator was placed
590 can affect the results. For this reason, the mastoid on which the bone vibrator was
591 placed shall be noted.

592 If no response occurs at the maximum output level of the audiometer, a downward
593 arrow should be drawn, attached to the corner of the appropriate symbol, see Figure 6.
594 These symbols should not be connected with the line to symbols representing measured
595 thresholds.

596 *Note:*

597 Some of the symbols used in audiological software packages may differ from those
598 recommended here. This is acceptable as long as the results are clear and unambiguous.

599







600
601
602

Name:
Date of birth:

Date:
Case No:

RIGHT

LEFT



...

	Right	Left
Air conduction, masked if necessary	○	×
Bone conduction, not masked		△
Bone conduction, masked	□]
Uncomfortable loudness level	└	┘

604
605
606
607
608
609

Audiometer type & serial number: _____
Earphone type: _____
Date of last objective calibration: _____
Tester: _____ Signature: _____
Comments: _____


610 **Figure 5 - Recommended format for audiogram form**





6013 Working audiograms

612 Working audiograms may be useful for some purposes, especially training, and they may
613 use shaded symbols for air conduction to indicate possible shadow points, which have
614 not been masked. Open symbols should be used to indicate the true hearing threshold,
615 which have been masked if necessary. Figure 6 is an example of a working audiogram.
616 The application of masking in the testing of the right a-c thresholds revealed shadows at
617 250 Hz and 500 Hz but not at 1000 Hz and 2000 Hz. (These latter two symbols could
618 have been half-filled in, indicating that masking had been performed, for training or
619 audit purposes.) The right ear threshold at 8000 Hz is greater than 120 dB HL, as
620 indicated by the arrow, and it should not be connected by line to the other results for
621 the right ear.



622
623 **Figure 6 - Illustration of a working audiogram.**

624

6014 Masking levels

626 Less experienced testers and students might find it useful to retain any masking charts,
627 or record the masking levels used, for training or audit purposes.

628





6095 Notes

630 If the tester has any doubts about the accuracy of any results, including any thresholds
631 where cross-hearing was indicated but masking not completed, these shall be noted.

632 The tester's name, signature and date of test should be noted on the audiogram form.
633 For electronic copies of the audiogram, the tester's name without signature is
634 acceptable. A note should also be made of the audiometer used, including the type of
635 earphones, and the date of the last objective calibration.

636 When a computerised audiometer is used, care must be taken to ensure all results are
637 recorded and stored correctly. In particular, some systems automatically delete not-
638 masked thresholds when masked thresholds are recorded, even though with bone-
639 conduction tests the initial not-masked result may correctly refer to the contralateral
640 ear. Testers should ensure all potentially useful data are retained.

14. Calibration

6421 Stage A: routine checking and subjective tests

643 In order to check the audiometer is functioning across the range, checks shall be carried
644 out by someone with sufficiently good hearing to detect any faults such as described
645 below. They should be carried out in the normal test room, with the equipment set up
646 as installed. These checks should be logged. Where apparent faults are noted,
647 equipment shall not be used until correct performance has been confirmed.

648 Tests 1 to 8 should be carried out daily.

- 649 1. Clean and examine the audiometer and all accessories. Check earphone cushions,
650 plugs, main leads and accessory leads for signs of wear or damage. Any badly worn
651 or damaged parts should be replaced. If any transducers are replaced, then the
652 audiometer must undergo a Stage B check.
- 653 2. Switch on equipment and leave for the recommended warm-up time (if no warm-up
654 period is quoted by the manufacturer, allow 5 minutes for circuits to stabilise). Carry
655 out any setting-up adjustments as specified by the manufacturer. On battery-
656 powered equipment, check battery state using the specified method. Check that
657 earphone and bone vibrator serial numbers tally with those on the instrument's





658 calibration certificate. An instrument's transducers shall not be changed unless a full
659 Stage B calibration is undertaken.

660 3. Check that the audiometer output is approximately correct on both a-c and b-c by
661 sweeping through at a hearing level of just audible tones (e.g. 10 dB HL or 15 dB HL).
662 This test should be performed at all appropriate frequencies and for both earphones
663 and the bone vibrator.

664 4. Check that the masking noise is approximately correct at all frequencies through
665 both earphones, at a level of 60 dB HL.

666 5. Perform a high-level listening check on a-c and b-c at all frequencies used, on all
667 appropriate functions and on both earphones (e.g. 60 dB HL for a-c, 40 dB HL for b-
668 c). Listen for proper functioning, absence of distortion, freedom from clicks when
669 presenting the tone etc.

670 6. Check all earphones and the bone vibrator for absence of distortion and
671 intermittency; check plugs and leads for intermittency.

672 7. Check that all the switches are secure and that lights and indicators work correctly.

673 8. Check that the subject response button works correctly.

674 Tests 9 to 12 should be carried out weekly.

675 9. Listen at low levels for any sign of noise or hum, for unwanted sounds or for any
676 change in tone quality as masking is introduced. Check that attenuators do
677 attenuate the signals over their full range and that attenuators which are intended
678 to be operated while a tone is being delivered are free from electrical or mechanical
679 noise. Check that interrupter keys operate silently and that no noise radiated from
680 the instrument is audible at the subject's position.

681 10. Check subject communication speech circuits.

682 11. Check tension of headset headband and bone vibrator headband. Ensure that swivel
683 joints are free to return without being excessively slack. Check headbands and swivel
684 joints for signs of wear strain or metal fatigue.

685 12. Perform an audiogram on a known subject, and check for significant deviation from
686 previous audiograms (e.g. 10 dB or greater).





6872 Stage B: periodic objective tests

688 Stage B checks are objective tests which ideally should be performed every 3 months,
689 although this period can be extended provided the Stage A checks are regularly and
690 carefully applied and it can be shown that the equipment is stable and reliable. The
691 maximum interval between checks should not exceed 12 months. They should
692 preferably be carried out in the normal test room, with the equipment set up as
693 installed, particularly if inter-connecting leads are used through a booth wall.

694 Measure and compare with the appropriate standards:

695 13. Frequencies of test signals

696 14. Sound pressure levels in an acoustic coupler or artificial ear from earphones

697 15. Vibratory force levels on a mechanical coupler from bone vibrators

698 16. Levels of masking noise

699 17. Attenuator steps over a significant part of the range

700 18. Harmonic distortion

1013 Stage C: basic calibration tests

702 Stage C checks need not be employed on a routine basis if Stage A and B checks are
703 regularly performed. They will only be required when a serious error or fault occurs, or
704 when, after a long period of time, it is suspected that the equipment may no longer be
705 performing fully to specifications. It may be advisable to submit equipment for a Stage C
706 check after, for example, five years' use if it has not received such a test in that time in
707 the course of repair.

708 Stage C checks should be such that after the audiometric equipment has been submitted
709 for a basic calibration, it shall meet the relevant requirements given in BS EN 60645-1. A
710 suggested minimum requirement for a Stage C check would include all items covered at
711 Stage B plus:

712 19. Rise and fall times of test tones

713 20. Interrupter effectiveness





- 714 21. Cross-talk between transducers and channels
- 715 22. Masking noise spectra
- 716 23. Distortion of speech and other external input systems

717 *Note:*

718 If insert earphones are used, separate measurements at all three stages must be made
719 for them. On some equipment it is possible to store two sets of calibration values,
720 however for others it may be necessary to use correction factors for the second set of
721 earphones.

12. References

- 723 British Society of Audiology Procedure for Processing Documents. British Society of
724 Audiology. www.thebsa.org.uk
- 725 British Society of Audiology (2004). Pure tone air and bone conduction threshold
726 audiometry with and without masking and determination of uncomfortable loudness
727 levels. British Society of Audiology.
- 728 British Society of Audiology Recommended Procedure: Ear examination. British Society
729 of Audiology. www.thebsa.org.uk
- 730 British Society of Audiology. Recommended Procedure: Determination of uncomfortable
731 loudness levels. British Society of Audiology www.thebsa.org.uk .
- 732 Bell I, Goodsell S, Thornton ARD (1980) A brief communication on bone conduction
733 artefacts. Br J Audiol 14: 73-75.
- 734 Boothroyd A, Cawkwell S. (1970) Vibrotactile thresholds in pure tone audiometry. Acta
735 Otolaryngol 69: 381-387.
- 736 Cooper J, Lightfoot G (2000) A modified pure tone audiometry technique for medico-
737 legal assessment. Br J Audiol 34: 37-45
- 738 Fagelson M, Martin FN (1994) Sound pressure in the external auditory canal during
739 bone-conduction testing. J Am Acad Audiol 5: 379-383





- 740 Harkrider AW, Martin FN (1998) Quantifying air-conducted acoustic radiation from the
741 bone-conduction vibrator. J Am Acad Audiol; 9: 410-416
- 742 The Health & Safety Executive (2005) Controlling Noise at Work. The Control of Noise at
743 Work Regulations 2005. Guidance on Regulations. Crown.
- 744 Lightfoot GR (1979) Air-borne radiation from bone conduction transducers. Br J Audiol
745 1979; 13: 53-56.
- 746 Lightfoot GR, Hughes JB (1993) Bone conduction errors at high frequencies: implications
747 for clinical and medico-legal practice. J Laryngol Otol 107: 305-308.
- 748 Lightfoot GR (2000) Audiometer calibration: interpreting and applying the standards. Br J
749 Audiol 34: 311-316
- 750 Margolis RH, Eikelboom RH, Johnson C, Ginter SM, Swanepoel DW, Moore BCJ (2013)
751 False air-bone gaps at 4 kHz in listeners with normal hearing and sensorineural hearing
752 loss. Int J Audiol 52: 526-532.
- 753 Munro KJ, Agnew N (1999) A comparison of inter-aural attenuation with the Etymotic
754 ER-3A insert earphone and the Telephonics TDH-39 supra-aural earphone. Br J Audiol
755 33: 259-262.
- 756 Shipton MS, John AJ, Robinson DW (1980) Air-radiated sound from bone vibration
757 transducers and its implications for bone conduction audiometry. Br J Audiol 14: 86-99.
- 758
- 759
- 760
- 761
- 762





Appendix A. Authors and acknowledgments

764 This revision was conducted by the BSA Professional Guidance Group between
765 September 2016 and August 2018 in accordance with the BSA Procedure for Processing
766 Documents. The Group thanks all involved with previous versions of this document and
767 all who contributed to this review including those who contributed during the
768 consultation in March 2017. An electronic copy of the anonymised comments (from 19
769 individuals) received during the most recent consultation, and the responses to these by
770 the authors, is available from BSA on request.

771 **Conflicts of interest** – none declared.

772

Appendix B. Standards relevant to audiometry

774 BS EN ISO 389-1: Acoustics. Reference Zero for the Calibration of Audiometric
775 Equipment. Part 1: Reference Equivalent Threshold Sound Pressure Levels for Pure
776 Tones and Supra-aural Earphones. (Identical to ISO 389-1)

777 BS EN ISO 389-2: Acoustics. Reference Zero for the Calibration of Audiometric
778 Equipment. Part 2: Reference Equivalent Threshold Sound Pressure Levels for Pure
779 Tones and Insert Earphones. (Identical to ISO 389-2)

780 BS EN ISO 389-3: Acoustics. Reference Zero for the Calibration of Audiometric
781 Equipment. Part 3: Reference Equivalent Threshold Sound Pressure Levels for Pure
782 Tones and Bone Vibrators. (Identical to ISO 389-3)

783 BS EN ISO 389-4: Acoustics. Reference Zero for the Calibration of Audiometric
784 Equipment. Part 4: Reference Levels for Narrow-band Masking Noise. (Identical to ISO
785 389-4)

786 BS EN ISO 389-4: Acoustics. Reference Zero for the Calibration of Audiometric
787 Equipment. Part 8: Reference Equivalent Threshold Sound Pressure Levels for Pure
788 Tones and Circumaural Earphones. (Identical to ISO 389-8)





- 789 BS EN ISO 7029:2000. Acoustics. Statistical Distribution of Hearing Thresholds as a
790 Function of Age. (Identical to ISO 7029:2000.)
- 791 BS EN ISO 8253-1: Acoustics. Audiometric Test Methods. Part 1: Basic Pure Tone Air and
792 Bone Conduction Threshold Audiometry. (Identical to ISO 8253-1)
- 793 BS EN 60645-1: Electroacoustics. Audiological Equipment. Part 1: pure-tone
794 audiometers. (Identical to IEC 60645-1)
- 795 Further information relevant to audiometric standards can be found on the National
796 Physical Laboratory website: www.npl.co.uk.
- 797

Appendix C. Permitted ambient noise levels for audiometry

- 799 To enable the accurate testing of normal air- and bone-conduction hearing threshold
800 levels down to 0 dB HL, ambient sound pressure levels should not exceed any of the
801 levels shown in Tables 1 and 2 respectively (from BS EN ISO 8253-1). To measure
802 minimum hearing threshold down to levels other than 0 dB HL, calculate the maximum
803 permissible ambient sound pressure levels by adding the minimum hearing threshold
804 level required to the values in Tables 1 and 2. For example, to measure down to 10 dB
805 HL, add 10 dB to all the values in the table.
- 806





807 **Table 1**

808 *Maximum permissible ambient sound pressure levels for measuring air-conduction audiometry*
809 *(supra-aural earphones) to a minimum hearing level of 0 dB HL between frequencies 250 Hz and*
810 *8000 Hz.*

Mid-frequency of one-third octave band (Hz)	dB re 20 μ Pa	Mid-frequency of one-third octave band (Hz)	dB re 20 μ Pa	Mid-frequency of one-third octave band (Hz)	dB re 20 μ Pa
31.5	66	250	19	2000	30
40	62	315	18	2500	32
50	57	400	18	3150	34
63	52	500	18	4000	36
80	48	630	18	5000	35
100	43	800	20	6300	34
125	39	1000	23	8000	33
160	30	1250	25		
200	20	1600	27		

811

812 *Notes on ambient noise:*

813 Insert earphones (e.g. Etymotic Research ER3 and ER5) and noise-excluding earphones
814 (e.g. Audiocups) will not require such stringent ambient noise levels as they reduce the
815 amount of ambient noise reaching the ears, if they are fitted correctly. However, full
816 details of the frequency-specific attenuation characteristics of these devices needs to be
817 considered, together with full details of the ambient noise, before tests can be carried
818 out in environments that exceed the noise levels listed above.

819

820





Table 2

Maximum permissible ambient sound pressure levels for measuring bone-conduction audiometry to a minimum hearing level of 0 dB HL between frequencies 250 Hz and 8000 Hz.

Mid-frequency of one-third octave band (Hz)	dB re 20 µPa	Mid-frequency of one-third octave band (Hz)	dB re 20 µPa	Mid-frequency of one-third octave band (Hz)	dB re 20 µPa
31.5	63	250	13	2000	8
40	56	315	11	2500	6
50	49	400	9	3150	4
63	44	500	8	4000	2
80	39	630	8	5000	4
100	35	800	7	6300	9
125	28	1000	7	8000	15
160	21	1250	7		
200	15	1600	8		

824

